

TEMPERATURE CONTROLS PTY LTD

ACN: 075 298 592 ABN: 76 075 298 592

7 YAMMA STREET

SEFTON, NSW 2162 AUSTRALIA

PH: (02) 9721 8644 FAX: (02) 9738 9339 EMAIL: sales@temperature.com.au



THERMOCOUPLES * RTD SENSORS * THERMOWELLS * EXTENSION CABLES * LEVEL SWITCHES

Ref 1038

CERTIFICATE OF TEST TYPE K THERMOCOUPLE

STABILISED THERMOCOUPLE

Stabilised: The thermocouple was annealed at 540⁰C for 24 hours

Test Procedure: The Type K thermocouple was checked against a GLA
Electronica Model 6500 Precision Calibrator
Serial No:

The Thermocouple was placed in a Isotech block calibrator
and checked at the following points.

Test Temperature:

Test Results: Type K Thermocouple Single 6.0 mm diameter MIMS

Tag Nos:

Comment: The Thermocouple was manufactured from Okazaki 6.0 mm
diameter MIMS Cable and is within standard grade limits of
the error.

Uncertainty: The estimated uncertainty of measurements is $\pm 1^{\circ}\text{C}$.

Date of Test:

Presented To:

Order No:

Signed

.....
Michelle Evans
Testing Officer