

TEMPERATURE CONTROLS PTY LTD

ACN: 075 298 592 ABN: 76 075 298 592

7 YAMMA STREET

SEFTON, NSW 2162 AUSTRALIA

PH: (02) 9721 8644 FAX: (02) 9738 9339 EMAIL: sales@temperature.com.au



THERMOCOUPLES * RTD SENSORS * THERMOWELLS * EXTENSION CABLES * LEVEL SWITCHES

Ref: 1039

CERTIFICATE OF COMPLIANCE **FROM THE ORIGINAL EQUIPMENT MANUFACTURER.**

Customer: Jord International PTY LTD

Date: 24/05/06

Country of Origin: The RTD Sensors were manufactured at our Sefton Temperature Sensor manufacturing facility, Sydney, Australia.

Instruments checked: PT100 Ω Class A RTD Sensors 3 wire

Quantity:

Accuracy: All RTD Sensors are filled with Class B Elements to DIN standard
EN 60751 1996 / IEC 751 1995
Tolerance @ 0 °C = 0.1% or 0.3 °C

Comment: This is to confirm that the temperature sensors supplied on the above complied with the order and specification required.
All assemblies were function-tested and visually inspected to ensure compliance.

Signed:

Phillip Lonsdale
Testing Officer