

# TEMPERATURE CONTROLS PTY LTD

ACN: 075 298 592 ABN: 76 075 298 592

SYDNEY: PH: (02) 9721 8644 FAX: (02) 9738 9339 EMAIL: [sales@temperature.com.au](mailto:sales@temperature.com.au)

MELBOURNE: PH: (03) 9384 0444 FAX: (03) 9384 0204 EMAIL: [vicsales@temperature.com.au](mailto:vicsales@temperature.com.au)



THEMOCOUPLES \* RTD SENSORS \* THERMOWELLS \* EXTENSION CABLES \* LEVEL SWITCHES

Ref: 1052

## CERTIFICATE OF TEST

### For Regal TC 305 K Digital Thermometer

**Procedure:** Calibrated against a Thermoelectric Cal Pal Serial No. 92080037, traceable to National Measurement Laboratory, United States National Institute of Standards via Phillips Defence Systems Pty Limited – Certificate No. R/N R986524.

**Test Results:** The instrument was checked for calibration at.

**Test Temperature:** 60.0 °C and 85.0 °C  
**Actual Temperature:** 59.9 °C and 85.1 °C

**Comments:** The instrument was found to conform manufacturers specifications.

**Instruments Tested:** **CIE TC 305K** **Serial No: 227528**

Accuracy of Instrument Tested  
 $\pm (0.3\% \text{ rdg} + 1^\circ\text{C}) -50^\circ\text{C to } 1000^\circ\text{C}$

**TC15K015030** **Serial No: 03070**

**Date of Test** 11/07/06

**Presented To:** Hobart Food Equipment

Testing Officer

Joe Pelzmann