

TEMPERATURE CONTROLS PTY LTD

ACN: 075 298 592 ABN: 76 075 298 592

7 YAMMA STREET

SEFTON, NSW 2162 AUSTRALIA

PH: (02) 9721 8644 FAX: (02) 9738 9339 EMAIL: sales@temperature.com.au



THERMOCOUPLES * RTD SENSORS * THERMOWELLS * EXTENSION CABLES * LEVEL SWITCHES

Saved: 1057

TEST DATA REPORT (TRANSMITTERS)

| Test Equipment | Serial # | Cal. Date | Due. Date | Test Transmitter | |
|----------------|-----------|-----------|-----------|-------------------------------------|---------------|
| | | / / | / / | Model # APAQ- HRF PT100 Transmitter | |
| GLA Calibrator | 009748652 | 19/07/06 | 19/07/06 | Job # YOKOGAWA AUSTRALIA | |
| | | / / | / / | Serial # 03113 | |
| | | / / | / / | Tech | Date 26/09/06 |

| Sensor Input | | | Range | |
|------------------|--------|--------------------|------------------------|--|
| 3 wire PT100 RTD | | | | |
| Input % | | Actual Output - mA | Calculated Output - mA | |
| (0) | 0 °C | 4.000 | 4.000mA | |
| | Ω | mA | | |
| (25) | 25°C | 8.006 | 8.000mA | |
| | | Ma | | |
| (50) | 50 °C | 12.005 | 12.000 mA | |
| | Ω | mA | | |
| (75) | 75 °C | 16.001 | 16.000Ma | |
| | | mA | | |
| (100) | 100 °C | 20.001 | 20.000mA | |
| | Ω | mA | | |

| | | | |
|---------------------|-----|----------------------------------|------------------------------------|
| Burnout | | <input type="checkbox"/> Upscale | <input type="checkbox"/> Downscale |
| Zero Adjustability | | <input type="checkbox"/> Accept | <input type="checkbox"/> Reject |
| Span Adjustability | | <input type="checkbox"/> Accept | <input type="checkbox"/> Reject |
| Line voltage effect | %1% | <input type="checkbox"/> Accept | <input type="checkbox"/> Reject |

Other tests: (Describe below)

Tested by

Date Tested

6/09/06

Phillip Lonsdale

MELBOURNE OFFICE: PHONE (03) 9384 0444 FAX: (03) 9384 0204
PROUDLY MANUFACTURING IN AUSTRALIA