

TEMPERATURE CONTROLS PTY LTD

ACN: 075 298 592 ABN: 76 075 298 592



7 YAMMA STREET SEFTON, NSW 2162 AUSTRALIA

PHONE: 61 2 9721 8644 FAX: 61 2 9738 9339

INTERNET: www.temperature.com.au EMAIL: sales@temperature.com.au

ISO9001:2000

Lic 14412

THERMOCOUPLES * RTD SENSORS * THERMOWELLS * EXTENSION CABLES * LEVEL SWITCHES

Ref: 1178

CALIBRATION CERTIFICATE 1/10 DIN RTD SENSORS

Procedure:

The 1/10 DIN RTD SENSORS

- Water bath – ambient to 100 °C checked connected to LABFACILITY BENCHTOP PRECISION INDICATOR S/n 04373 The measurement results reported

Test Temperature:

30 °C 60 °C & 85 °C

Test Results:

The instrument was checked for calibration at the above temperatures, and is accurate to within ± 0.1 °C

Instruments Tested:

RTD Hand Probe

Serial No: 04370

Accuracy of Instrument Tested

@ 30 °C reading 29.95 °C

@ 60 °C reading 59.95 °C

@ 85 °C reading 85.03 °C

Date of Test

17/06/06

Presented To:

HEASTERN INDUSTRIES

Order No.

60H92

Testing Officer

Michael Donnelly