


Special Accuracy (roll certificate required)  
 @ 300°C +/- 0.40% ( or +/- 1.0 °C)

Roll certificate required - Must be special limits

Special Accuracy  
 300°C +/- 0.40% ( or +/- 1.0 °C)  
 TYRE TEMPERATURE THERMOCOUPLE  
 TYPE K, J, T OR PT100  
 HAND HELD IMMERSION PROBE

<b>TEMPERATURE CONTROLS PTY LTD</b>		Certificate No: QEC 14412 ISO 9001:2000
<b>CONDITIONS</b> THIS DRAWING IS OUR EXCLUSIVE PROPERTY AND MUST NOT BE COPIED OR HANDED TO ANY OTHER PARTIES WITH OUT WRITTEN PERMISSION. IT SHALL BE RETURNED UPON REQUEST	SYDNEY : 7 YAMMA ST SEFTON , NSW 2162 PHONE : (02) 9721 8644 FAX : (02) 9738 9339 MELBOURNE : 8/280 Whitehall St YARRAVILLE VIC 3013 PHONE : (03) 9687 0000 FAX : (03) 9687 1900 EMAIL : sales@temperature.com.au	
	TITLE 	
SIZE : A4	DATE : 07/10/10	<b>DWG NO : SD 085</b>
SCALE : NTS	DRAWING BY R.B	<b>REV : 01</b>