



Instrument Connection	D2	D1	F	U
1/2"NPT or BSP	38.1mm	19mm	45mm	Specify
1/2"NPT or BSP	28.5mm	14mm	45mm	Specify

**NOTE: Standard taper length up to 400mm
for longer taper lengths or full taper refer to SD 036FT**

Specify	Quality assurances Documentation	Certificate Reference
	Material certificate – barstock	ORIGINAL
	Hydrostatic Pressure test certificates - Internal	1043
	NACE - Heat treatment certificate	1045
	<i>Cleaned for oxygen IEC 60877 Sealed in plastic bag</i>	<i>1055</i>
	Thermowell calculation to ASME PTC 19.3	1135
	Tagging	STAMP
	ITP (specify)	1056, D
	Other	Specify

THERMOWELL WELD IN TYPE

TEMPERATURE CONTROLS PTY LTD

QUALITY ENDORSED COMPANY
Certificate No: QEC 14412
AS/NZS 9001:2008

CONDITIONS

THIS DRAWING IS OUR
EXCLUSIVE PROPERTY
AND MUST NOT BE
COPIED OR HANDED
TO ANY OTHER PARTIES
WITH OUT WRITTEN
PERMISSION. IT SHALL
BE RETURNED UPON
REQUEST

SYDNEY: 7 YAMMA ST SEFTON , NSW 2162
PHONE: (02) 9721 8644 FAX : (02) 9738 9339
MELBOURNE: 8/280 Whitehall St YARRAVILLE, VIC 3013
PHONE: (03) 9687 0000 FAX : (03) 9687 1900
BRISBANE: PO Box 5138 Alexandra Hills, QLD 4161
PHONE: 0418 222 868
EMAIL: sales@temperature.com.au

TITLE

SIZE : A4 DATE :20/02/13 DWG NO :SD 200

SCALE : NTS DRAWING BY R.B REV : 02